



Attestation of Conformity

No. E8A 076845 0003 Rev. 00

Holder of Certificate: SENKO Co., Ltd.

73, Oesammi-ro, 15beon-gil Osan-si, Gyeonggi-do 18111 REPUBLIC OF KOREA

Name of Object: Electrical equ. for measurement, control

and laboratory use

(Sampling type Gas Detector)

Model(s): SI-H100, SI-H100(NH3)

Description of

Object:

Rated input voltage:

24 V d.c.

Rated input current: 0.2 A

Tested according to:

EN 50270:2015

EN 61000-6-4:2007/A1:2011

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.:

DREKCEE1905-0323

Date.

2019-05-16

(Byung-Soo Kang)